

T E S T R E P O R T

KOTITI No : 8220-1509-100319-001

Issue No : 5373041552

Receipt Date : 23 / 03 / 2020

Issue Date : 07 / 04 / 2020

Sample Name	MNS Maskfilter	Test Object	Research/Reference/Etc
Lot No.	MN200316	MDF/MFG	-
Weight/Volume	-	Expiry Date	16 / 03 / 2023
Company	WooreeMNS Co., Ltd.	President	Sung Young Bin
Address	79, Seounggok-ro, Danwon-gu, Ansan-si, Gyeonggi-do, Korea	Telephone	82-31-508-6060

Test Item & Test Result

Test Item	Test Standards	Unit	Result	Test Method
Pigment	no color	-	no color	KQC
acid or alkali	no red color	-	no red color	KQC
Fluorescent	no fluorescent	-	no fluorescent	KQC
Formaldehyde	not be darker than potassium chromate	-	not be darker than potassium chromate	KQC
Filtering efficiency(NaCl)	94 ≧	%	Before pretreatment : 99/99/99 After pretreatment : 99/99/99	MFDS Guideline
Filtering efficiency(Paraffin oil)	94 ≧	%	Before pretreatment : 98/98/98 After pretreatment : 98/98/97	MFDS Guideline

Researcher : Yoon Sub Cho

Technical Manager : Man Sik Wang

- ※ The test results above shall be only subject to the requested test items.
- ※ In the case of insufficient space on the page, test items and test results may be written in the appendix.
- ※ When advertising test results or indicating test results on containers or packages, the entire content of the test report shall be indicated.
- ※ This report produced in relation to the services will not be related with MFDS. And MFDS is not legally liable.

KOTITI Testing & Research Institute



Designated as Cosmetics Inspection Institute(Ministry of Food and Drug Safety, No.14)